

## Packing Adapter R11

49331-PA2

Part No.

**Packing Adapter, PEEK**, for 4.6 mm ID columns and 7.5 mm ID slurry reservoirs, ARE-Applied Research brand. Bulk Price.



### Additional Info

This PEEK column packing adapter is used when packing ARE brand 4.6 mm ID PEEK columns (*usually longer than 100 mm in length*) and using our 7.5 mm ID, larger packing reservoirs for sufficient volume of packing material during the packing process.

This adapter is made from PEEK which ensures a biocompatibility from the beginning of the packing and column process.

Threads are 1/2 - 20 (for 4.6 mm columns) on one side and 5/8 - 18 (for 7.5 mm columns)) to connect the packing reservoir on the other side. Sealing rings are required: 1 each [49330-PA2](#) and 2 each [49330-PA4](#).

*NOTE: This adapter is designed for packing only All-PEEK HPLC column hardware and not recommended for use with stainless steel packing systems.*

### MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

Printed On 2024/05/20