# **Product Sheet**



## Catalog Number

### 58045-V13-K

Syringe Filters, 13mm, PVDF, 0.45um Pore Size. Red Polypropylene, AQ Brand. 1000/CS.



### Additional Info

Advanced Quality Polyvinylidene Fluoride (PVDF) Syringe Filters, AQ<sup>™</sup> brand, for Pharmaceutical Grade HPLC & Dissolution Testing. Manufactured with a unique 0.45µm PVDF membrane in a 13mm polypropylene housing, these syringe filters are made for laboratories that require consistent, high performance results with ease of use especially during high volume work. Excellent choice when laboratories cannot afford poor chromatographic results

Validated for many regulatory requirements, they are designed, tested and QC certified. They comply with US FDA CFR Title 21, Part 210.3(b)(6) for minimum Fiber Release and with USP <788> test for particulate matter in injectables as well as USP for "Sterile Water for Injection" and therefore do not require any initial flushing or discarding of the first 200µl of filtrate saving time and avoiding investigation of failed methods. The polypropylene plastic housing material complies with the requirements of US FDA CFR Title 21 Part 177.1520 are heat sealed to prevent leaching and pass a test for oxidizable matter as per the USP.

#### MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451 tel. (732) 380-8900, fax (910) 769-9435 Email: customers@mtc-usa.com Website: <u>www.mtc-usa.com</u>



Printed On 2024/05/18