

Capsaicinoids in Hot Sauce Analysis by HPLC - Tips and Suggestions

If you have a hot sauce extract containing nordihydrocapsaicin, capsaicin, dihydrocapsaicin, and homodihydrocapsaicin, the Cogent Bidentate C18 2.0™ HPLC column is a good choice for sufficient separation when the compounds are analyzed by LC-MS; look for the [M+H]⁺ ions.


The following gradient should be a good starting point:

Mobile Phase: A: DI H₂O / 0.1% formic acid (v/v)

B: Methanol

Gradient:	time (min.)	%B
	0	10
	1	10
	9	70
	10	70
	11	10

You may get EICs which look like this:



1. nordihydrocapsaicin, 294.2064 m/z [M+H]⁺
2. capsaicin, 306.2064 m/z [M+H]⁺
3. dihydrocapsaicin, 308.2220 m/z [M+H]⁺
4. homodihydrocapsaicin, 322.2377 m/z [M+H]⁺

For peaks 1 and 2, you will be able to obtain separate peaks for each compound in the EICs so there will not be an issue of interference from co-elution. **In terms of sample preparation, try using a Soxhlet extraction procedure.**

Click [HERE](#) for Cogent Bidentate C18™ ordering information



Printed from the Chrom Resource Center
Copyright 2025, All Rights Apply
MicroSolv Technology Corporation
9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900
Fax: (910) 769-9435
Email: customers@mtc-usa.com
Website: www.mtc-usa.com