



What is the maximum concentration of ammonium fluoride to use in a mobile phase to avoid column damage - Tips & Suggestions

If you want to use ammonium fluoride as a mobile phase additive to increase signal to noise in LCMS, the recommended limit is about 2 mM for TYPE-C Columns.

Our studies have shown that concentrations of 5 mM can cause some damage to Cogent TYPE-C columns and therefore should be avoided. Therefore we suggest no more than 1 to 2 mM ammonium fluoride.

It is known that the damaging effects to the column are due to the presence of fluoride rather than the ammonium because we have been able to use ammonium acetate, for instance up to 16.5 mM.



Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com