

## Isopropanol Rarely Used as a Solvent in HPLC or LCMS Methods - HPLC Primer

*Date: 14-APRIL-2020 Last Updated: 8-NOVEMBER-2025*

While isopropanol (IPA) is a common laboratory solvent, it is infrequently used as a primary mobile phase component in HPLC or LC-MS methods. Here's why:

---

### 1. Higher Viscosity = Higher Backpressure

IPA is significantly more viscous than commonly used solvents like acetonitrile (ACN) or methanol (MeOH). This increased viscosity leads to:

- Higher system backpressure
- Reduced flow rate flexibility
- Potential strain on pumps and fittings

As a result, IPA is less practical for routine use, especially in high-throughput or high-pressure systems.

---

### 2. Limited Selectivity Advantages

In terms of chromatographic selectivity, IPA offers few unique benefits. When visualized in a solvent selectivity triangle (with ACN, MeOH, and tetrahydrofuran [THF] at the corners), IPA typically falls within the range achievable by blending these other solvents. Therefore, most desired selectivity effects can be replicated without using IPA.

---

### 3. Where IPA Is Useful

Despite its limitations, IPA has valuable niche applications in HPLC workflows:

- Solvent bridging: IPA is miscible with both aqueous and non-aqueous solvents, making it ideal for transitioning between reversed-phase and normal-phase methods.
  - Column cleaning: IPA is effective at removing strongly retained hydrophobic contaminants. It can be used:
    - As a wash step at the end of a gradient
    - In dedicated column cleaning protocols, including backflushing
- 

 Need help selecting the right solvent for your method? Our technical team is here to assist.

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)