



Triton X-100 Assay by HPLC For Cleaning Methods - AppNote

A Sensitive Assay for Detection of a Non-ionic Surfactant

Triton X-100 is used in the sample purification steps of various materials as well as for cleaning apparatus in pharmaceutical manufacturing. Due to the toxic nature of Triton X-100, a sensitive method for its detection and quantification is very important to identify any residual amounts.

This simple Isocratic Method may be used to verify residual amounts of this Surfactant have been removed from final material or from vital manufacturing equipment.



Peak: Triton X-100

Method Conditions:

Column: Cogent Bidentate C18™, 4 μm, 100 Å

Catalog No.: 40018-75P

Dimensions: 4.6 x 75 mm

Mobile Phase: 80:20 Acetonitrile / DI Water with 0.1% Formic Acid

Injection vol.: 5 μL

Flow rate: 1.0 mL / minute

Detection: UV @ 234 nm

Sample Preparation: Triton X-100, 50 ug / mL in DI Water

Note: Triton X-100 is one of the most widely used nonionic surfactants for lysing cells to extract protein and other cellular organelles. Triton X-100 ($C_{14}H_{22}O(C_2H_4O)_n$) has a hydrophilic polyethylene oxide chain and an aromatic hydrocarbon lipophilic or hydrophobic group. The hydrocarbon group is a 4-phenyl group.

This Method was developed by and is presented courtesy of [ARL- Eutech Scientific Services, Inc.](#)



Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com