



Hydrocortisone Assay Alternate L1 Column – AppNote

Alternate L1 column used for a similar USP assay method of Hydrocortisone

This Similar USP Hydrocortisone Assay is easily performed and the five injection overlay below demonstrates Run to Run consistency and Precision. The Peak Efficiency is very good and Peak Tailing guidelines are easily met with RSD values of 0.5%.

This demonstrates a good alternate Column for your Hydrocortisone USP Method.



Peak: Hydrocortisone

Method Conditions:

Column: Cogent MicroBee C18™, 5 um, 125 Å

Catalog No.: [59305-03](#)

Dimensions: 3.9 x 300 mm

Mobile Phase: 25:75 Acetonitrile / DI Water

Injection vol.: 15 μ L
Flow rate: 1.0 mL / minute
Detection: UV @ 254 nm

Sample Preparation: Hydrocortisone 50 μ g / mL in Methanol

Note 1: Hydrocortisone, a Corticosteroid, and is similar to a natural hormone produced by your Adrenal Glands. Hydrocortisone is used to treat immune, inflammatory, and neoplastic conditions.

Note 2: The Cogent MicroBee C18™ performs very similarly to the Waters® μ Bondapak C18®.



Printed from the Chrom Resource Center
Copyright 2025, All Rights Apply
MicroSolv Technology Corporation
9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900
Fax: (910) 769-9435
Email: customers@mtc-usa.com
Website: www.mtc-usa.com