



Can I use chlorinated or fluorinated solvents with PEEK frits and columns for HPLC - FAQ

Yes!

PEEK is compatible with most halogenated organic solvents that are used in HPLC.

ARE-APPLIED RESEARCH™

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com