

What is the maximum back pressure in PEEK FlexiTubes - FAQ

Date: 22-DECEMBER-2023 Last Updated: 18-OCTOBER-2025

It's not possible to specify a definitive maximum back pressure for PEEK FlexiTubes due to the wide range of variables that influence it.

Back pressure is affected by several factors, including:

- **Flow rate**
- **Solvent density**
- **Temperature**
- **Tubing dimensions**
- **System configuration**

Because these parameters can vary significantly depending on the application, users should evaluate back pressure under their specific operating conditions.

ARE-APPLIED RESEARCH™

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com