


Zinc EDTA Complex Analyzed by LC-MS - AppNote

No Ion Pair Agents Needed

Using conventional analytical methods, retention of metal-EDTA complexes is accomplished using Ion Pair Reversed Phase Chromatography. However, the Ion Pair Agents used in the Mobile Phase are not compatible with Mass Spectrometry.

In this AppNote, only Formic Acid is needed in the Mobile Phase in order to obtain retention of a Zinc-EDTA complex. The figure shows an EIC of the analyte spiked in a soil extract matrix.



Peak: Zn-EDTA complex 354.7 m/z

Method Conditions

Column: Cogent Diamond Hydride™, 4 μm, 100 Å

Catalog No.: [70000-15P-2](#)

Dimensions: 2.1 x 150 mm

Mobile Phase:

- A: DI Water / 0.1% Formic Acid (v/v)
- B: Acetonitrile / 0.1% Formic Acid (v/v)

Gradient:

Time (minutes)	%B
0	90
5	20
8	20
9	90

Post Time: 2 minutes**Flow rate:** 0.4 mL / minute**Detection:** ESI – NEG - Agilent 6210 MSD TOF Mass Spectrometer**Injection vol.:** 1 µL

Sample Preparation: A Soil sample was spiked with Zn-EDTA complex at a level of 2000 µm. After extraction with DI Water (shaking for 24 hours), the sample was filtered using a 0.45 µm Nylon Syringe Filter (MICROSOLV Tech Corp.) and diluted with Acetonitrile 1:10 before injection.

t_o: 0.9 minutes

Note: EDTA (Ethylenediaminetetracetic Acid) has been used to promote metal uptake from contaminated Soil. As such, the ESI-MS technique described here can be used for the determination of Zn-EDTA complexes in Soil samples.



Attachment No 173 Zinc-EDTA Complex pdf 0.3 Mb [Download File](#)

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.comWebsite: www.mtc-usa.com