



Are Septa in Autosampler Vial Caps Self Sealing Inert and Compatible with Organic Solvents - FAQ

Date: 18-JULY-2016 Last Updated: 8-NOVEMBER-2025

All of your non-slit septa are self-sealing and lined with Polypropylene, PTFE or other fluoropolymer for inertness.

HPLC and GC solvents do not present a compatibility issue but acetonitrile and ethers are very volatile and may evaporate after a single use or piercing of the septa. You may need to replace the cap after an injection if you need extreme control over evaporation from the vial.

Click [HERE](#) for screw cap ordering information and pictures

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com