



Can I Sterilize Screw Caps for Autosampler Vials that Have Silicone Rubber with PTFE Septa - FAQ

Date: 7-JUNE-2012 Last Updated: 27-OCOTBER-2025

Yes, if you use heat / steam or gamma irradiation, the AQR, AQ, MICROSOLV and the BASIK brands of 8mm, 9mm and 10mm screw caps can be sterilized or autoclaved for up to 15 minutes.

Click [HERE](#) for ordering information and pictures of autosampler vials

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com