



Have the Plastic Amber Autosampler Vials Changed Color Compared to Earlier Batches - FAQ

Date: 29-APRIL-2015 Last Updated: 27-OCTOBER-2025

The amber vials catalog # *9502S-PP-AMBER* have been discontinued due to instabilities in the material which was causing color changes. These items have been replaced with catalog # [9502S-PP-AM](#). You may notice the newer vials are more opaque.

[Click HERE for 9502S-PP-AM Ordering Information](#)

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com