

## What is the Main Difference Between Standard and LCMS Compatible Polypropylene Plastic Autosampler Vials - FAQ

*Date: 13-APRIL-2018 Last Updated: 27-OCTOBER-2025*

The differences between standard plastic microvials like [9532S-0PV](#) and the LCMS compatible microvials [9532S-MS](#) are that item 9532S-MS is made from a special type of polypropylene and we do not use "mold release agents" in the molding process making it less likely to have any extractables or leachables under any circumstance.

*Acetonitrile is likely to remove existing mold release agents which can show up in chromatograms as spurious peaks.*

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)