



Vials Listed as 12 x 32 mm Compatibility with Autosamplers' Maximum Diameter of 11.8 mm - HPLC Primer

Date: 13-SEPTEMBER-2013 Last Updated: 27-OCTOBER-2025

12 x 32mm OD, 2ml vials are listed as 12mm diameter for standardization purposes within the industry. This is a nominal value only (name only) and the exact value is suitable for use with your autosampler.

The outer diameter is actually 11.8mm for most 12x32mm vials.

Click [HERE](#) for autosampler vial ordering information and pictures.

AUTOSAMPLER
VIALS AND CAPS

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com