

Avoid Loss of Concentration of Heparin When Using AQ Brand Syringe Filters - Tips & Suggestions

Date: 24-JANUARY-2014 Last Updated: 8-NOVEMBER-2025

The most appropriate membrane type to use when filtering heparin containing samples is AQ™ brand regenerated cellulose (RC) syringe filter.

Heparin has strongly acidic sulfonate groups (negatively charged) that will bind with nylon filter membranes. PTFE is not suitable because of its hydrophobicity and passes water through it is not possible.

Click [HERE](#) for ordering information and pictures of AQ brand of syringe filters.



25mm Regenerated Cellulose Filter

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com