



Will I See any Differences in my Chromatography if I Use AQ Brand Syringe Filters - FAQ

Date: 22-APRIL-2012 Last Updated: 3-NOVEMBER-2025

If you use proper filtering techniques and a proper filter you should not see any differences in your chromatography, CE or UV testing.

Syringe filters only protect your equipment and columns and never participate in separations or SPE.

Click [HERE](#) for syringe filters ordering information and pictures.

Attachment

MICROSOLV filters equivalency study pdf [Download File](#)

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com