

Serum Samples Use Suggestion for Cogent Diamond Hydride Columns in LCMS Methods - Tips & Suggestions

Date: 12-DECEMBER-2012 Last Updated: 30-NOVEMBER-2025

When using serum and other biological samples, to keep the columns clean, the A Solvent with DI water and methanol or IPA should be used continuously and not only for conditioning.

We recommend with this mobile phase:

- Solvent A: 50:50 DI water / methanol (or 2-propanol) + 0.1% formic acid (for positive mode MS) or +10 mM ammonium acetate or formate (for negative mode MS).
- Solvent B: The same as you would ordinarily use with Aqueous Normal Phase ANP for example, 98% Acetonitrile and 2% DI water with 1-5mM acid or base.

Cleaning Tip: Methanol or 2-propanol (IPA) helps to clean the column between injections, which is necessary for analysis of complex matrices such as plasma, serum, urine, etc. If for any reason only DI water (with additives) has to be used as solvent A, you can add a washing step in the autosampler sequence every 10 injections using the methanol or 2-propanol-containing solvent A listed above.

Related article: [General solvent conditioning for TYPE-C Silica columns](#)



Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com