



Stability of UDC Cholesterol HPLC Column as a Stationary Phase - Tech Information

While cholesterol is a large molecule, the UDC-Cholesterol phase has proven to be stable in 0.1% v/v TFA and 0.1% v/v H₃PO₄ at acid pH's and up through pH 8.00 with 20mM NH₄COOH buffer with NH₄OH.

A pH in excess of pH 8.00 should be avoided as this phase has a mono (single point) attachment to the silica matrix and an intermediate ester link in its structure.

For higher pH's the Cogent Bidentate C18™ is recommended.



Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com