



Can I use Silica C columns in a Normal Phase gradient method - FAQ

Yes!

Although this is not easy with ordinary silica columns due to solvent de-mixing & water uptake, the Cogent Silica-C™ column is not limited by this and is easy to use even with fast gradients.



Printed from the Chrom Resource Center
Copyright 2025, All Rights Apply
MicroSolv Technology Corporation
9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900
Fax: (910) 769-9435
Email: customers@mtc-usa.com
Website: www.mtc-usa.com