




Tizanidine HCl Tablet Precision by HPLC - AppNote

Excellent Run to Run Precision

Tizanidine has several amine groups which can present problems to chromatographers. Residual silanols on the surface of ordinary Silica-based Columns can electrostatically interact with these groups, causing peak tailing.

The Cogent Amide Column, used in the Application Note is virtually free of residual silanols and therefore a symmetrical peak shape can be obtained. An overlay of five consecutive runs demonstrates the run- to-run precision when using these columns.



Tizanidine

Peak:

Tizanidine

Method Conditions

Column: Cogent Amide™, 4µm, 100Å

Catalog No.: [40036-05P](#)

Dimensions: 4.6 x 50mm

Mobile Phase:

A: 90% DI Water / 10% Acetonitrile / 0.1% Formic Acid (v/v)

B: Acetonitrile / 0.1% Formic Acid (v/v)

Gradient:

Time (Minutes)	%B
0	90
2	90
6	40
7	90

Flow rate: 1.0 mL/minute

Detection: UV @ 230nm

Injection vol.: 3µL

Sample Preparation:

4 mg strength Tizanidine HCl tablet was ground and weighed in a 10 mL volumetric flask. A portion of 50/50 solvent A/ solvent B diluent was added and the flask was sonicated 10 minutes. It was then diluted to mark and filtered with a 0.45µm Nylon syringe filter (MICROSOLV Tech Corp). An aliquot was then diluted 1:10 with a diluent of 10% solvent A / 90%.

Note: Tizanidine is a centrally acting α 2- adrenergic agonist used to treat spasms, cramping, tightness of muscles, and related conditions. It is available under the trade name Zanaflex® as well as generic versions.



Attachment No 356 Tizanidine Amide.pdf 0.5 Mb [Download File](#)

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com