

Can I Use Benzene as a Mobile Phase Solvent in HPLC - FAQ

Date: 13-OCTOBER-2014 Last Updated: 3-NOVEMBER-2025

Benzene can be used as an HPLC solvent... in theory.

It is compatible with standard HPLC instrumentation and will not damage an HPLC column.

However, **we would not recommend** using this solvent for two main reasons. First, it is strongly UV-absorbing so would not be suitable for UV-based detection methods. Second, it is a known carcinogen and therefore less safe than many other solvents which are available. If you need something with similar properties, we would suggest toluene. It is often used as a benzene substitute and is less toxic.

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com