

Can I use the UV-vis Calibration Kit to qualify detection at 200 nm - FAQ

Date: 15-OCTOBER-2020 Last Updated: 5-JULY-2025

The UV-Vis Calibration Kit includes two key standards:

- **Caffeine Standard:** Provides absorbance peaks at **205 nm** and **273 nm**.
- **Holmium Oxide Solution:** Offers 14 well-defined absorbance bands spanning 241 nm to 641 nm.

While the kit does not include a direct reference point at 200 nm, the caffeine peak at 205 nm allows for extrapolation to this wavelength. This is typically sufficient for qualification purposes, as 200 nm is near the lower detection limit for most HPLC solvents due to their UV cutoff.

Note: Always verify that your solvent system does not interfere at this wavelength.

 Click [HERE](#) for HSQ™ & PQ Kits™ ordering information and images.

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com