

## Validation of the Method in Qualification Kits Is Not Necessary - Tech Information

DATE: 17-FEBRUARY-2017 Last Updated: 5-JULY-2025

Validation of the Chemical Solutions brand of HPLC Qualification Kits™ method is not required. This method is specifically designed to qualify the performance of the HPLC instrument itself.

For example, if you're considering validation parameters like LOD (Limit of Detection) or LOQ (Limit of Quantitation), these are not applicable in this context. The HPLC Qualification Kits™ assumes the detector is functioning correctly—but verifying that assumption is exactly what the kit is designed to do. In other words, the method relies on the detector's performance, which the kit is actively qualifying.

Additionally, there is no federal regulatory requirement mandating validation of this method, since it is not intended for product testing or regulatory submission.

---

 **Important Note:** Before performing any method validation on your HPLC system, you must first complete IQ/OQ/PQ (Installation, Operational, and Performance Qualification) to ensure the instrument is functioning within specifications.

---

 Click [HERE](#) for PQ & HSQ Kit™ ordering information and product images.

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)