

## Mobile Phase in the Chemical Solutions HPLC Qualification Kits - Tech Information

Date: 7-APRIL-2012 Last Updated: 7-JULY-2025

The "Do It Yourself", **Chemical Solutions HPLC Qualification Kits** are designed for ease of use and compatibility with most standard HPLC systems. The mobile phase formulation is optimized to support both **isocratic** and **gradient** system performance checks using NIST-traceable standards.

### Mobile Phase Composition

- 14:86 Acetonitrile : 0.1% Acetic Acid in Water (v/v)
  - To prepare the aqueous component: Add 1 mL of glacial acetic acid per liter of water to make a 0.1% solution.
  - Mix 140 mL of acetonitrile with 860 mL of the 0.1% acetic acid solution to make 1 liter of mobile phase.

*Tolerance: 12%-16% in acetonitrile absolute composition.*

*0.08%-0.12% acetic acid absolute composition.*

### This composition provides:

- Low UV background, ideal for UV detector qualification
- Stable baseline for both isocratic and gradient testing
- Compatibility with a wide range of analytes and column chemistries

### Volume Requirements

- For isocratic system testing: Prepare 1 liter of mobile phase.
- For gradient system testing: Prepare 2 liters to ensure sufficient volume for method development, equilibration, and multiple injections.

### Additional Tips

- Always use HPLC-grade solvents and freshly prepare the mobile phase to avoid contamination or degradation.
- Filter and degas the mobile phase before use to prevent baseline noise and pump cavitation.
- Store unused mobile phase in amber bottles at room temperature and use within 24–48 hours for best results.
- 

 Click [HERE](#) for PQ and HSQ Kit™ ordering information and images.

Printed from the Chrom Resource Center

Copyright 2025, All Rights Apply

**MicroSolv Technology Corporation**

9158 Industrial Blvd. NE, Leland, NC 28451

Tel: (732) 380-8900

Fax: (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)