

Catalog Number

49330-F2

Frit, PEEK, 2 μ m, encased in a PEEK ring, 0.20 inch OD .102 inch ID, 0.062 inch thick, for use with 2.1mm HPLC columns. EPP-Porous Products Brand, 1. EA.



Additional Info

These all PEEK frits are made for use for the ARE-Applied Research brand, all PEEK, 2.1 mm ID HPLC column hardware.

Dimensions are: *0.20 inch (1.57 mm) thick and 0.102 inch (2.59 mm) diameter. These frits have a 2 μ m porosity and the PEEK ring retains biocompatibility.*

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Price

\$40.45

Printed On 2024/05/02