

Catalog Number

49330-PA2

Packing Seal, PEEK, for use with 4.6 mm ID column packing adapters and 4.6 mm or 7.5 mm ID slurry reservoirs. ARE-Applied Research brand. 1 EA.



Additional Info

This "packing" seal is for when using the packing adapter [49330-PA1](#) to pack 4.6mm ID PEEK columns with length greater than 100 mm requiring with a 7.5 mm ID slurry reservoir [49333-SR3](#). Normally this seal is used with 2 each [49333-PS packing seals](#) to properly connect the slurry reservoir to the 4.6 mm column being packed. This seal reduces the flow from the 7.5 mm ID reservoir to the desired 4.6 mm ID column.

It is also used when using the packing adapter to pack 4.6 mm ID PEEK column with length equal to or less than 100 mm.

This Seal is made from PEEK which ensures a biocompatibility from the beginning of the packing and column process.

NOTE: This sealing ring is designed for packing only All-PEEK HPLC column hardware and not recommended for use with stainless steel packing systems.

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Price

\$10.59

Printed On 2024/04/29