

Catalog Number

49330-PA5

Packing Adapter, PEEK, for 2.1 mm ID columns and 4.6 mm ID slurry reservoirs, ARE-Applied Research brand. 1 EA.



Additional Info

This Packing Adapter is made for all ARE brand 2.1 mm ID PEEK columns when using our 4.6 mm ID packing reservoir. This adapter is designed for packing only All-PEEK HPLC column hardware and not recommended for use with stainless steel packing systems.

This adapter is made from PEEK which ensures a biocompatibility from the beginning of the packing and column process.

MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: customers@mtc-usa.com

Website: www.mtc-usa.com

Price

\$59.70

Printed On 2024/02/25