

Catalog Number

49911-01

Ferrules, Replacement, for Di-Ad, Direct Adaptive HPLC Column Connectors. PEEK. MicroSolv Brand. 10/PK.



Additional Info

After several hundred connections to HPLC columns or to the instruments, it is advisable to replace the "all PEEK Ferrule" of a Direct Adaptive (Di-Ad™) HPLC column connector, to insure good connections and prevent leaking during use. It is easy and simple to remove and insert this replacement.

Click [HERE](#) for instructions on how to replace ferrules on Di-Ad™ Column Connectors.

MicroSolv Technology Corporation
9158 Industrial Blvd. NE, Leland, NC 28451
tel. (732) 380-8900, fax (910) 769-9435
Email: customers@mtc-usa.com
Website: www.mtc-usa.com

Price

\$230.61

Printed On 2024/06/20